

Deutsche Akkreditierungsstelle GmbH

Entrusted according to Section 8 subsection 1 AkkStelleG in connection with Section 1 subsection 1 AkkStelleGBV
Signatory to the Multilateral Agreements of EA, ILAC and IAF for Mutual Recognition

Accreditation



The Deutsche Akkreditierungsstelle GmbH attests that the calibration laboratory

EMH Energie-Messtechnik GmbH
Vor dem Hassel 2, 21438 Brackel

is competent under the terms of DIN EN ISO/IEC 17025:2018 to carry out calibrations in the following fields:

Electrical Quantities

DC and Low Frequency Quantities

- DC voltage
- DC current
- AC voltage
- AC current
- Electric energy
- Electric power
- Current ratio

The accreditation certificate shall only apply in connection with the notice of accreditation of 02.08.2022 with the accreditation number D-K-12011-01. It comprises the cover sheet, the reverse side of the cover sheet and the following annex with a total of 3 pages.

Registration number of the certificate: **D-K -12011-01-00**



Berlin,
02.08.2022

Dr. Florian Witt
Head of Technical Unit

Translation issued:
02.08.2022

Technical Unit

The certificate together with its annex reflects the status at the time of the date of issue. The current status of the scope of accreditation can be found in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH.
<https://www.dakks.de/en/content/accredited-bodies-dakks>

This document is a translation. The definitive version is the original German accreditation certificate.
See notes overleaf.

Deutsche Akkreditierungsstelle GmbH

Office Berlin
Spittelmarkt 10
10117 Berlin

Office Frankfurt am Main
Europa-Allee 52
60327 Frankfurt am Main

Office Braunschweig
Bundesallee 100
38116 Braunschweig

The publication of extracts of the accreditation certificate is subject to the prior written approval by Deutsche Akkreditierungsstelle GmbH (DAkkS). Exempted is the unchanged form of separate disseminations of the cover sheet by the conformity assessment body mentioned overleaf.

No impression shall be made that the accreditation also extends to fields beyond the scope of accreditation attested by DAkkS.

The accreditation was granted pursuant to the Act on the Accreditation Body (AkkStelleG) of 31 July 2009 (Federal Law Gazette I p. 2625) and the Regulation (EC) No 765/2008 of the European Parliament and of the Council of 9 July 2008 setting out the requirements for accreditation and market surveillance relating to the marketing of products (Official Journal of the European Union L 218 of 9 July 2008, p. 30). DAkkS is a signatory to the Multilateral Agreements for Mutual Recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Cooperation (ILAC). The signatories to these agreements recognise each other's accreditations.

The up-to-date state of membership can be retrieved from the following websites:

EA: www.european-accreditation.org

ILAC: www.ilac.org

IAF: www.iaf.nu

Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D-K-12011-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 02.08.2022

Date of issue: 02.08.2022

Holder of certificate:

EMH Energie-Messtechnik GmbH
Vor dem Hassel 2, 21438 Brackel

Calibration in the fields:

Electrical Quantities

DC and Low Frequency Quantities

- **DC voltage**
- **DC current**
- **AC voltage**
- **AC current**
- **Electric energy**
- **Electric power**
- **Current ratio**

The management system requirements in DIN EN ISO/IEC 17025 are written in language relevant to operations of calibration laboratories and operate generally in accordance with the principles of DIN EN ISO 9001.

*The certificate together with its annex reflects the status at the time of the date of issue. The current status of the scope of accreditation can be found in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH.
<https://www.dakks.de/en/content/accredited-bodies-dakks>*

Annex to the accreditation certificate D-K-12011-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
AC voltage	60V, 120V, 240V, 480V	45 Hz $\leq f \leq$ 65 Hz	44 $\cdot 10^{-6}$	Comparison with HEG K2005 <i>f: Frequency</i>
	30 V to 480 V		50 $\cdot 10^{-6}$	
	25 mA		51 $\cdot 10^{-6}$	
	50 mA, 100 mA, 250 mA, 500 mA, 1 A, 2,5 A, 5 A, 10 A		38 $\cdot 10^{-6}$	
	25 A, 50 A, 100 A		44 $\cdot 10^{-6}$	
	20 mA to < 50 mA		57 $\cdot 10^{-6}$	
	50 mA to < 10 A		45 $\cdot 10^{-6}$	
	10 A to 100 A		50 $\cdot 10^{-6}$	
AC voltage	30 V to 480 V	45 Hz $\leq f \leq$ 65 Hz	64 $\cdot 10^{-6}$	Comparison with EMH K2006 <i>f: Frequency</i>
AC current	5 mA to 10 mA		0,24 $\cdot 10^{-3}$	
	> 10 mA to 20 mA		0,17 $\cdot 10^{-3}$	
	> 20 mA to 50 mA		0,1 $\cdot 10^{-3}$	
	> 50 mA to 160 A		70 $\cdot 10^{-6}$	
DC voltage	100 mV to 1000 V		17 $\cdot 10^{-6}$	Comparison with digital multimeter FLUKE 8588A Comparison with digital multimeter FLUKE 8588A and current transformer Danisense DS400ID with transducer Danisense DSSIU-4-1U
DC current	10 mA to 1 A		68 $\cdot 10^{-6}$	
	> 1 A to 100 A		72 $\cdot 10^{-6}$	
	> 100 A to 600 A		0,23 $\cdot 10^{-3}$	
AC active power and energy	750 mW to 4800 W 750 mWh to 4800 Wh	45 Hz $\leq f \leq$ 65 Hz 0,25 $\leq \cos\varphi \leq$ 1 60 V, 120 V, 240 V, 480 V 50 mA, 100 mA, 250 mA, 500 mA, 1 A, 2,5 A, 5 A, 10 A	51 $\cdot 10^{-6}$	Comparison with HEG K2005 Relative measurement uncertainty related to the apparent power or energy
	375 W to 48 kW 375 Wh to 48 kWh	45 Hz $\leq f \leq$ 65 Hz 0,25 $\leq \cos\varphi \leq$ 1 60 V, 120 V, 240 V, 480 V 25 A, 50 A, 100 A	60 $\cdot 10^{-6}$	
	150 mW to < 26 W 150 mWh to < 26 Wh	45 Hz $\leq f \leq$ 65 Hz 0,25 $\leq \cos\varphi \leq$ 1 30 V to 480 V 20 mA to < 50 mA	0,15 $\cdot 10^{-3}$	
	375 mW to < 4,8 kW 375 mWh to < 4,8 kWh	45 Hz $\leq f \leq$ 65 Hz 0,25 $\leq \cos\varphi \leq$ 1 30 V to 480 V 50 mA to < 10 A	57 $\cdot 10^{-6}$	
	> 75 W to 4,8 kW > 75 Wh to 4,8 kWh	45 Hz $\leq f \leq$ 65 Hz 0,25 $\leq \cos\varphi \leq$ 1 30 V to 480 V 10 A to 100 A	64 $\cdot 10^{-6}$	
	37,5 mW to 2,4 W 37,5 mWh to 2,4 Wh 37,5 mVar to 2,4 VAr 37,5 mVArh to 2,4 VArh 150 mVA to 2,4 VA 150 mVAh to 2,4 VAh	45 Hz $\leq f \leq$ 65 Hz $\pm 0,25 \leq \cos\varphi \leq \pm 1$ $\pm 0,25 \leq \sin\varphi \leq \pm 1$ 30 V to 240 V 5 mA to 10 mA	0,24 $\cdot 10^{-3}$	

Date of issue: 02.08.2022

Valid from: 02.08.2022

Page 2 of 3

Annex to the accreditation certificate D-K-12011-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range		Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
	> 75 mW	to 4,8 W	45 Hz $\leq f \leq$ 65 Hz	$0,2 \cdot 10^{-3}$	
	> 75 mWh	to 4,8 Wh	$\pm 0,25 \leq \cos\varphi \leq \pm 1$		
	> 75 mVAr	to 4,8 VAr	$\pm 0,25 \leq \sin\varphi \leq \pm 1$		
	> 75 mVArh	to 4,8 VArh	30 V to 240 V		
	> 300 mVA	to 4,8 VA	$> 10 \text{ mA to } 20 \text{ mA}$		
	> 300 mVAh	to 4,8 VAh			
	> 150 mW	to 12 W	45 Hz $\leq f \leq$ 65 Hz	$0,12 \cdot 10^{-3}$	
	> 150 mWh	to 12 Wh	$\pm 0,25 \leq \cos\varphi \leq \pm 1$		
	> 150 mVAr	to 12 VAr	$\pm 0,25 \leq \sin\varphi \leq \pm 1$		
	> 150 mVArh	to 12 VArh	30 V to 240 V		
	> 600 mVA	to 12 VA	$> 20 \text{ mA to } 50 \text{ mA}$		
	> 600 mVAh	to 12 VAh			
	> 375 mW	to 9,6 kW	45 Hz $\leq f \leq$ 65 Hz	$92 \cdot 10^{-6}$	
	> 375 mWh	to 9,6 kWh	$\pm 0,25 \leq \cos\varphi \leq \pm 1$		
	> 375 mVAr	to 9,6 kVAr	$\pm 0,25 \leq \sin\varphi \leq \pm 1$		
	> 375 mVArh	to 9,6 kVArh	30 V to 480 V		
	> 1,5 VA	to 9,6 kVA	$> 50 \text{ mA to } 20 \text{ A}$		
	> 1,5 VAh	to 9,6 kVAh			
	> 150 W	to 76,8 kW	45 Hz $\leq f \leq$ 65 Hz	$92 \cdot 10^{-6}$	
	> 150 Wh	to 76,8 kWh	$\pm 0,25 \leq \cos\varphi \leq \pm 1$		
	> 150 VAr	to 76,8 kVAr	$\pm 0,25 \leq \sin\varphi \leq \pm 1$		
	> 150 VArh	to 76,8 kVArh	30 V to 480 V		
	> 600 VA	to 76,8 kVA	$> 20 \text{ A to } 160 \text{ A}$		
	> 600 VAh	to 76,8 kVAh			
DC electric power and electric energy	1 mW	to 1 kW	100 mV to 1000 V	$70 \cdot 10^{-6}$	Comparison with digital multimeter FLUKE 8588A and current transformer Danisense DS400ID with transducer Danisense DSSIU-4-1U
	1 mWh	to 1 kWh	$10 \text{ mA to } 1 \text{ A}$		
	100 mW	to 100 kW	100 mV to 1000 V	$74 \cdot 10^{-6}$	
Current transformers with transformation ratio 1:1	100 mWh	to 100 kWh	$> 1 \text{ A to } 100 \text{ A}$		Substitution method with two reference standards
	10 W	to 600 kW	100 mV to 1000 V	$0,23 \cdot 10^{-3}$	
	10 Wh	to 600 kWh	$> 100 \text{ A to } 600 \text{ A}$		
	20 mA	to < 50 mA	45 Hz $\leq f \leq$ 65 Hz	$0,12 \cdot 10^{-3}$ $0,47' \triangleq 0,014 \text{ crad}$	
	50 mA	to 160 A		$0,11 \cdot 10^{-3}$ $0,41' \triangleq 0,012 \text{ crad}$	

Abbreviations used:

CMC	Calibration and measurement capabilities (Kalibrier- und Messmöglichkeiten)
EMH	EMH Energie-Messtechnik GmbH
HEG	Hamburger Elektronik Gesellschaft

Date of issue: 02.08.2022

Valid from: 02.08.2022

Page 3 of 3